

PLATINUM LEVEL ONLY



2525 Tollgate Rd, Quakertown, PA 18951
Phone: 215-536-4400 Fax: 215-536-4096
www.precisionolutionsinc.com

Certificate of Calibration

Certificate #: 4080-22394
Service Order: 3123
Service Date: MM/DD/YYYY
Technician: Tech Name

Client: Customer Name
123 Street Road
Anywhere, USA

Location of Calibration: Customer Name
123 Street Road
Anywhere, USA

Manufacturer: Avery Weigh-Tronix
Model: Bench Scale
Serial #: 83743289
Capacity: 200 lb
Graduation: 1 lb
Cust Equip Desc: RID # ABC Rental Unit

Condition of Unit: Very Good
In Tolerance (As Left): Yes
Temperature (F): 75
Humidity (%): 32
Procedure*: Handbook 44

Test Data Unit of Measure: lb

* If "Handbook 44" then N.1.1 and Table 6 were utilized

Test Type	Weight Applied	As Found	Initial Error	As Left	Final Error
Increasing Load	10	10	0	10	0
Increasing Load	50	50	0	50	0
Increasing Load	100	100	0	100	0
Increasing Load	150	150	0	150	0
Increasing Load	200	200	0	200	0
Shift 1	75	75	0	75	0
Shift 2	75	75	0	75	0
Shift 3	75	75	0	75	0
Shift 4	75	75	0	75	0

Reference Standard
Set 6

Test #
#####

Reference standards used are traceable to the International System of Units (SI) through NIST

The uncertainty of measurement associated with the measurement results reported in this certificate is available upon request and was accounted for, where applicable, in making the decision of compliance or non-compliance according to the Test Method identified above.

Notes: Inspected, tested and calibrated.



PJLA Calibration
Accreditation# 59161

This device has been calibrated under a system that is accredited to ISO/IEC 17025:2005. This certificate may not be reproduced except in full, without written approval of Precision Solutions, Inc.

Authorized Signature

Daniel R. Baldwin, Service Manager