

**PLATINUM LEVEL ONLY**



2525 Tollgate Rd, Quakertown, PA 18951  
Phone: 215-536-4400 Fax: 215-536-4096  
www.precisionolutionsinc.com

**Certificate of Calibration**

**Certificate #:** 4080-22394  
**Service Order:** 3123  
**Service Date:** MM/DD/YYYY  
**Technician:** Tech Name

**Client:** Customer Name  
123 Street Road  
Anywhere, USA

**Location of Calibration:** Customer Name  
123 Street Road  
Anywhere, USA

**Manufacturer:** Avery Weigh-Tronix  
**Model:** Bench Scale  
**Serial #:** 83743289  
**Capacity:** 200 lb  
**Graduation:** 1 lb  
**Cust Equip Desc:** RID # ABC Rental Unit

**Condition of Unit:** Very Good  
**In Tolerance (As Left):** Yes  
**Temperature (F):** 75  
**Humidity (%):** 32  
**Procedure\*:** Handbook 44

**Test Data Unit of Measure:** lb

\* If "Handbook 44" then N.1.1 and Table 6 were utilized

Test Type	Weight Applied	As Found	Initial Error	As Left	Final Error
Increasing Load	10	10	0	10	0
Increasing Load	50	50	0	50	0
Increasing Load	100	100	0	100	0
Increasing Load	150	150	0	150	0
Increasing Load	200	200	0	200	0
Shift 1	75	75	0	75	0
Shift 2	75	75	0	75	0
Shift 3	75	75	0	75	0
Shift 4	75	75	0	75	0

**Reference Standard**  
Set 6

**Test #**  
#####

Reference standards used are traceable to the International System of Units (SI) through NIST

The uncertainty of measurement associated with the measurement results reported in this certificate is available upon request and was accounted for, where applicable, in making the decision of compliance or non-compliance according to the Test Method identified above.

Notes: Inspected, tested and calibrated.



Accreditation# 59161

Form # 237.6 Rev C

This device has been calibrated under a system that is accredited to ISO/IEC 17025:2005. This certificate may not be reproduced except in full, without written approval of Precision Solutions, Inc.

Authorized Signature

Daniel R. Baldwin, Service Manager