



2525 Tollgate Rd, Quakertown, PA 18951
 Phone: 215-536-4400 Fax: 215-536-4096
 www.precisionolutionsinc.com

Certificate of Inspection

Order Number: 38643-001271
Certificate Date: 17-Aug-2021
Technician: Brandyn Wismer
Approved By: Ryan Leatherman

Customer Information

Client: Customer Name
 123 Street Road
 AnyTown, PA 12345

Reference Standard(s) Used

ID	NIST Test #
Set #7	20946

The units listed have been tested with Precision Solutions, Inc. standards which are traceable to the International System of Units (SI) through NIST. The results reported are at the time of testing. Calibration was performed in accordance with Precision Solutions, Inc.'s procedures. This document shall not be reproduced, except in full, without the written approval of Precision Solutions, Inc. Unless otherwise specified by the customer, the units under test were determined to be in or out of tolerance based upon the test measurement being within the simple acceptance range inclusive of lower and upper tolerance limits. Tolerance decisions are not further modified using guard bands or uncertainty.

#	Make	Model	Serial #	As Found	As Left	Service Date
1	Uline	H 5836	02311042005	PASS	PASS	04-Aug-2021
2	Avery Weigh-Tronix	WI-130	10602	PASS	PASS	04-Aug-2021
3	Avery Weigh-Tronix	ZM-303	122350201	FAIL	PASS	04-Aug-2021
4	Avery Weigh-Tronix	ZM-301	161650044	FAIL	PASS	04-Aug-2021
6	Avery Weigh-Tronix	WI-127	23482	FAIL	PASS	04-Aug-2021
7	Avery Weigh-Tronix	WI-127	24199	PASS	PASS	04-Aug-2021
8	Ohaus	T51XW	B403232693	PASS	PASS	04-Aug-2021

Comments: