



2525 Tollgate Rd, Quakertown, PA 18951
Phone: 215-536-4400 Fax: 215-536-4096
www.precisionolutionsinc.com

Electronic Data Report

Report #: 38643-001268-04

Service Date: 05-Aug-2021

Technician: Trent Bradford

Customer Information

Client: Customer Name Inc
123 Street Road
Quakertown, PA 18951

Location of Calibration: Customer Name Inc
123 Street Road
Quakertown, PA 18951

Asset Information

Manufacturer: Ohaus
Model: ARC120
Serial #: J04351213
Customer ID: QC
Location: QC
Range:
Capacity: 3100 g
Resolution: 0.01 g

As Found Readings

Applied	Reading
Zero (g)	
0.00	0.00
Increasing Load (g)	
100.00	99.98
300.00	299.93
1000.00	999.85
2000.00	1999.74
3000.00	2999.77
Shift Round (g)	
500.00	499.89
500.00	499.87
500.00	499.91

As Left Readings

Applied	Reading
Zero (g)	
0.00	0.00
Increasing Load (g)	
100.00	100.00
300.00	300.00
1000.00	1000.00
2000.00	1999.95
3000.00	3000.00
Shift Round (g)	
500.00	499.99
500.00	499.99
500.00	500.00

Reference Standard(s) Used ID:

Set #4

Comments:

Inspected, tested, and calibrated